

Certificate of Analysis

Product Name:

Standard Colour Solution YG (Yellowish Green) for examination of the colour of solutions, PurCert (according to Chinese Pharmacopoeia)

Product No.: PSR50451
Lot No.: Sample COA
Brand: PureSynth
CAS No.: NA
Molecular Formula: NA
Molecular Weight: NA
Date of Mfg.: NA
Date of Exp.: NA

Assignment of CIE Tristimulus and L*a*b* values (as per ChP Method 3):

Test Results:	X ₁₀ = 83.454	L* = 96.8567
	Y ₁₀ = 92.089	a* = -7.3744
	Z ₁₀ = 62.893	b* = 27.2419

Tristimulus Test Method:

The result reported above was determined by analysis of a sample of this lot taken at the time of manufacture.

The spectral characteristics of the sample were recorded in a 10 mm cell between 380nm and 780nm using a Shimadzu UV1800 spectrophotometer in accordance with an in-house test method. The tristimulus and L*a*b* values reported above were calculated from this data according to the ChP and ASTM method E308, using Illuminant D65 and the CIE 1964 Supplementary Standard (10°) Observer.

Specification and Analysis of ChP colour solution:

Test results for the Standard Solutions used in the manufacture of this product were within the specifications outlined below:

Standard Solutions	Specified Conc. (g/l)
Cobaltous chloride	59.5 ±0.1
Potassium dichromate	0.8 ±0.1
Cupric sulfate	62.4 ±0.1

The concentration values were assigned using a compleximetric titration method that is traceable to NIST SRM 915b or 136e.

The Certificate of Analysis (COA) is valid only for the lot specified and within the product's shelf life. Retesting is recommended after the expiry date or if storage conditions are compromised. We guarantee a proper quality within our General Conditions of Sales. The conditions are subject to change without prior notice.

For queries on this COA, please contact our technical team at technical@pure-synth.com

Colour solution formulation according to ChP:

Table 1: Composition of stock Reference Solution

Quantity Standard Solution (ml) per 100 ml of Reference Solution YG

Cobaltous chloride	Potassium dichromate	Cupric sulfate	Water
1.2	22.8	7.2	68.8

Table 2: Composition of Reference Solutions 0.5 – 10

No. of Reference solution	0.5	1	2	3	4	5
Stock Reference solution (ml)	0.3	0.5	1.0	1.5	2.0	2.5
Water (ml)	9.8	9.5	9.0	8.5	8.0	7.5

No. of Reference solution	6	7	8	9	10
Stock Reference solution (ml)	3.0	4.5	6.0	7.5	10.0
Water (ml)	7.0	5.5	4.0	2.5	0.0

Remarks:

This is to certify that the mentioned batch has been analysed at the time of quality release and has been found to be in compliance with prescribed specifications

The current Specification sheet may be available at www.pure-synth.com

Inspect on receipt

PureSynth Research Chemicals GmbH

This is electronically-generated copy & is valid without a signature

The Certificate of Analysis (COA) is valid only for the lot specified and within the product's shelf life. Retesting is recommended after the expiry date or if storage conditions are compromised. We guarantee a proper quality within our General Conditions of Sales. The conditions are subject to change without prior notice.

For queries on this COA, please contact our technical team at technical@pure-synth.com